



Certificate of Compliance

Certificate: 70022352

Master Contract: 263136

Project: 80002551

Date Issued: 2019-06-04

Issued to: APM Technologies(Dong Guan) Co.,Ltd
201 #7, Link Information, Industry Park,
Shuilianshan Road
Dongguan, Guangdong 523000
CHINA

Attention: Mr. Wen Cai

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by: *Jerry Ning*
Jerry Ning

PRODUCTS

CLASS 8721 06 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE (Certified to U.S. Standards)

Programmable DC Power Supply, Class I, Pollution Degree 2, permanently connected.

	Models	Input	Output
Part A	SP32VDC1600W	100-240 Vac, 50 – 60 Hz, 2000 VA	0 – 32 Vdc, 0 – 50 A, 1600 W
	SP40VDC1600W		0 – 40 Vdc, 0 – 40 A, 1600 W
	SP32VDC1000W	100-240 Vac, 50 – 60 Hz, 1300 VA	0 – 32 Vdc, 0 – 50 A, 1000 W
	SP40VDC1000W		0 – 40 Vdc, 0 – 40 A, 1000 W



Certificate: 70022352
Project: 80002551

Master Contract: 263136
Date Issued: 2019-06-04

	SP32VDC1200W	100-240 Vac, 50 – 60 Hz, 1500 VA	0 – 32 Vdc, 0 – 50 A, 1200 W
	SP40VDC1200W		0 – 40 Vdc, 0 – 40 A, 1200 W
Part B	SP60VDC1500W	100-240 Vac, 50 – 60 Hz, 1900 VA	0 – 60 Vdc, 0 – 25 A, 1500 W
	SP75VDC1500W		0 – 75 Vdc, 0 – 20 A, 1500 W Or 0 – 75 Vdc, 0 – 25 A, 1500 W
	SP60VDC1000W	100-240 Vac, 50 – 60 Hz, 1300 VA	0 – 60 Vdc, 0 – 25 A, 1000 W
	SP75VDC1000W		0 – 75 Vdc, 0 – 20 A, 1000 W Or 0 – 75 Vdc, 0 – 25 A, 1000 W
	SP60VDC1200W	100-240 Vac, 50 – 60 Hz, 1500 VA	0 – 60 Vdc, 0 – 25 A, 1200 W
	SP75VDC1200W		0 – 75 Vdc, 0 – 20 A, 1200 W Or 0 – 75 Vdc, 0 – 25 A, 1200 W
	SP75VDC600W	100-240 Vac, 50 – 60 Hz, 750 VA	0 – 75 Vdc, 0 – 25 A, 600 W

Notes:

1. Mode of operation: Continuous.
2. Environmental Conditions: Extended: 0 to 40 deg C, 2000 m max, 10% - 90% RH as specified by manufacturer.
3. The subject equipments are only to be installed by manufacturer trained personnel.



Certificate: 70022352

Master Contract: 263136

Project: 80002551

Date Issued: 2019-06-04

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 70022352

Master Contract: 263136

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70022352	2015-06-25	Original Certification.
80002551	2019-06-04	Update Report 70022352 for report corrections and additional of alternative rating for Models SP75VDC1000W, SP75VDC1200W and SP75VDC1500W. Add additional model SP75VDC600W.