



Certificate of Compliance

Certificate: 70177787

Master Contract: 263136

Project: 70177787

Date Issued: 2018-11-16

Issued to: **APM Technologies(Dong Guan) Co.,Ltd**
201 #7, Link Information,
Industry Park, Shuilianshan Road
Dongguan, Guangdong 523000
CHINA
Attention: Wen Cai

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US (indicating that products have been manufactured to the requirements of both Canadian and US Standards) or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by: *Jerry Ning*
Jerry Ning

PRODUCTS

CLASS 8721 06 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE (Certified to U.S. Standards)

Programmable AC Source, Class I, Pollution Degree 2, single phase, permanently connected.

Models	Input	Output
SP300VAC1500W	110-240 Vac, 50/60 Hz, 1800 VA	0-300 Vac, 0-13.8 A, 0-1500 VA 0-424 Vdc, 0-9.76 A, 0-1500 VA DC, 15Hz-1.2kHz
SP300VAC1000W	100-240 Vac, 50/60 Hz, 1200 VA	0-300 Vac, 0-9.2 A, 0-1000 VA 0-424 Vdc, 0-6.5 A, 0-1000 VA



Certificate: 70177787

Master Contract: 263136

Project: 70177787

Date Issued: 2018-11-16

		DC, 15Hz-1.2kHz
SP300VAC600W	100-240 Vac, 50/60 Hz, 800 VA	0-300 Vac, 0-5.6 A, 0-600 VA 0-424 Vdc, 0-3.96 A, 0-600 VA DC, 15Hz-1.2kHz

Notes:

1. Mode of operation: Continuous.
2. Environmental Conditions: Extended: 0 to 40 deg C, 2000 m max, 5% - 95% RH non-condensing as specified by manufacturer.
3. Equipment is only to be installed by manufacturer trained personnel.
4. Equipment is investigated for usage on 240 V single phase system of supply.

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 70177787

Master Contract: 263136

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70177787	2018-11-16	Original Certification.