

8 Cootamundra Cres.
Blackburn VIC 3130
AUSTRALIA
+61 (0)430 412 380
joshua@anzcompliance.com.au



Certificate No.: R23078

RCM CERTIFICATE

This certificate is issued to confirm that ANZ Electrical Compliance has certified the equipment described herein have comply with AS/NZS 4417.1 and can be marked with the Regulatory Compliance Mark.

Certificate Holder: APM Technologies (Dongguan) Co., Ltd.
Room 101, #1, Liaochaheng 1st Road, Liangbian
Village, Liaobu Town, Dongguan City, Guangdong
P.R. CHINA

Product: Programmable AC Power Supply

Trade Name / Manufacturer: APM Technologies (Dongguan) Co., Ltd.

Model(s): SP300VAC5000W, SP300VAC2000W, SP300VAC3000W,
SP300VAC4000W

Ratings: Input: 208-240 Va.c., 50/60 Hz
Output: 0-300 Va.c., 15 Hz – 1.2 kHz; 0-424 Vd.c.
SP300VAC2000W: Input: 2300 VA; Output: 0-2000 VA
SP300VAC3000W: Input: 3600 VA; Output: 0-3000 VA
SP300VAC4000W: Input: 4700 VA; Output: 0-4000 VA
SP300VAC5000W: Input: 6000 VA; Output: 0-5000 VA

Compliance Details: See below.

EESS: Not in-scope (Not for Household use)
Registered as Level 1 Equipment
Comply with AS 61010.1:2003
Test Report# CB 263136-80139084

ACMA/RSM: SDoC under Responsible Supplier E7663
Comply with IEC 61326-1:2020
Test Report Ref# 80139086

Issue Date: 13 March 2023



ANZ Electrical Compliance